## CERTIFICATE OF ANALYSIS BASELINE<sup>®</sup> Perchloric Acid

1A	PRODUCT NUMBER: S020201						LOT NUMBER: 2217010 AS					SAY (HCIO <sub>4</sub> , w/w): 67%						
3 Li < 2 11 Na	i 4 Be   < 0.05 of three aliquots subsampled from three samples representative of the lot. The samples are slowly evaporated to dryness. The resulting residue is reconstituted in a small volume of SEASTAR <sup>™</sup> BASELINE <sup>®</sup> 2% Nitric Acid / 2% Hydrogen Peroxide. For volatile elements (indicated by *), the acid samples are diluted then														7A			
< 10 19 K < 5	< 2 20 Ca < 100	3B	ue is subtra 4B <b>22 Ti</b> < 5	5B	<u>6B</u>	7B <b>25 Mn</b> < 0.5	<b>26 Fe</b> < 50	8 27 Co < 0.1	<mark>28 Ni</mark> < 10	1B <b>29 Cu</b> < 2	2B 30 Zn < 5	< 20 31 Ga < 1		<b>33 As</b> < 100				
<b>37 Rb</b> < 0.1	<b>38 Sr</b> < 0.1	<b>39 Y</b> < 0.02	<b>40 Zr</b> < 0.2	<b>41 Nb</b> < 10	<b>42 Mo</b> < 0.05			<b>45 Rh</b> < 0.05	<b>46 Pd</b> < 1	<b>47 Ag</b> < <mark>0.2</mark>	<mark>48 Cd</mark> < 0.1	<b>49 In</b> < 0.01	<mark>50 S</mark> n < 2	51 Sb < 1	<b>52 Te</b> < 0.1			
<b>55 Cs</b> < 0.01	<b>56 B</b> a < 0.5	<b>57 La</b> < 0.05	72 Hf < 1	<b>73 Та</b> < 20	74 W < 1				<b>78 Pt</b> < 10			<mark>81 TI</mark> < 0.01	82 Pb < 0.2	83 Bi < 0.01				

## ALL VALUES ARE REPORTED IN PARTS PER TRILLION (PPT)

KEY(1) Atomic Number(1) (2)(2) Elemental Symbol(3)(3) Concentration (m	< 0.1	<b>59 Pr</b> < 0.05	<b>60 Nd</b> < 0.05	<b>62 Sm</b> < 0.01	<b>63 Eu</b> < 0.01	<b>64 Gd</b> < 0.01	<b>65 Tb</b> < 0.01	66 Dy < 0.01	67 Ho < 0.01	68 Er < 0.01	<b>70 Yb</b> < 0.01	<b>71 Lu</b> < 0.01
(4) in ppt) (4) 1 Standard Devia (N=3)	ion 90 Th < 0.01	E.	<b>92 U</b> < 0.01									

HCIO<sub>4</sub> (65 - 71%): Properties Molar Mass: 100.46g/mol Density: 1.67 g/ml Molarity: 12 moles/litre Normality: 12 moles/litre

Bregternen



Greg Henson QA & RA Manager Release Date: Expiry Date: June 09, 2017 June 09, 2020

**)))(()** SEASTAR CHEMICALS



## SEASTAR<sup>™</sup>'s Product Integrity Guidelines:

We have found our products, unopened and sealed, maintain the certified integrity, or product quality, for their stated certification period under the following conditions:

- Store at room temperature, maximum range 15°C (59°F) to 25°C (77°F).
- Avoid exposure to sunlight or ultraviolet light sources.
- Open in a 'particle free' environment. SEASTAR recommends a HEPA or ULPA particle filtered trace metal clean room. Open product should be handled under Class 100 or ISO 5 clean room or better conditions.

Once opened, product integrity will depend on proper handling and exposure to contaminants. To reduce trace metal contamination, the inner pack of plastic bags and bottle should be opened under Class 100 or ISO 5 clean room or better conditions to maintain the integrity of the product. The use of plastic gloves, hair net and a clean room suit is also advised.

## Safety Guidelines:

PRIOR to opening or storing this product be sure to consult the Safety Data Sheet (SDS) to ensure safe storage and handling with regards to this hazardous material. This information must be read and understood prior to use or storage.

SAFETY HANDLING NOTES: Consult the SDS PRIOR to handling this product. Use proper safety apparel according to the recommendations of the SDS. Exposure controls and personal protection should include: a properly functioning fume hood, protection for eyes (safety glasses), hands (chemically compatible gloves), feet (chemically compatible boots), and exposed skin (splash protection and a chemically compatible apron). All of these items must conform to local/regional/national regulatory requirements.

Bregterin

Greg Henson QA & RA Manager

2061 Henry Avenue West, Sidney, BC, Canada V8L5Z6 Phone: 1-250-655-5880 Fax: 1-250-655-5888 Toll free: 1-800-663-2330 (North America only) Email: <u>seastar.technicalsupport@seastarchemicals.com</u> Web: <u>www.seastarchemicals.com</u>