

C E R T I F I C A T E O F A N A L Y S I S

BASELINE® Water

1A		PRODUCT NUMBER: S020901										LOT NUMBER: 9209090																					
		2A		<p>Most elements are determined by high resolution ICP-MS using sample preconcentration. The results are an average of three aliquots subsampled from three samples representative of the lot. The samples are slowly evaporated to dryness. The resulting residue is reconstituted in a small volume of SEASTAR™ BASELINE® 2% Nitric Acid / 2% Hydrogen Peroxide. Operations are conducted under Class 100 or better clean-room conditions. For volatile elements (indicated by *), the acid samples are diluted then directly injected into the ICP-MS. Values below 3 times the standard deviation of the blank are shown with '<', no blank value is subtracted.</p> <div>3B4B5B6B7B81B2B</div>										3A		4A	5A	6A	7A														
3	Li	4	Be											5	B																		
< 1		< 1												< 20																			
11	Na	12	Mg											13	Al																		
< 10		< 5												< 10																			
19	K	20	Ca	21	Sc	22	Ti	23	V	24	Cr	25	Mn	26	Fe	27	Co	28	Ni	29	Cu	30	Zn	31	Ga	32	Ge	33	As*	34	Se*		
< 10		< 10		< 1		< 5		< 1		< 10		< 1		< 10		< 1		< 10		< 5		< 5		< 1		< 1		< 10		< 50			
37	Rb	38	Sr	39	Y	40	Zr	41	Nb	42	Mo			44	Ru	45	Rh	46	Pd	47	Ag	48	Cd	49	In	50	Sn	51	Sb	52	Te		
< 1		< 1		< 1		< 1		< 1		< 1				< 1		< 1		< 5		< 5		< 1		< 1		< 5		< 10		< 1			
55	Cs	56	Ba	57	La	72	Hf	73	Ta	74	W	75	Re					78	Pt	79	Au	80	Hg*	81	Tl	82	Pb	83	Bi				
< 0.1		< 1		< 0.1		< 0.1		< 5		< 5		< 1						< 1		< 10		< 5		< 0.1		< 1		< 0.1					
ALL VALUES ARE REPORTED IN PARTS PER TRILLION (PPT)																																	

ALL VALUES ARE REPORTED IN PARTS PER TRILLION (PPT)

KEY	(1)	Atomic Number	58	Ce	59	Pr	60	Nd		62	Sm	63	Eu	64	Gd	65	Tb	66	Dy	67	Ho	68	Er	69	Tm	70	Yb	71	Lu	
	(1) (2)	(2) Elemental Symbol	< 0.1		< 0.1		< 0.1			< 0.1		< 0.1		< 0.1		< 0.1		< 0.1		< 0.1		< 0.1		< 0.1		< 0.1		< 0.1		
	(3)	(3) Concentration (mean in ppt)																												
	(4)	(4) 1 Standard Deviation (N=3)																												
			90	Th			92	U																						
			< 0.1				< 0.1																							



Additional Tests	Maximum Specification	Actual Value
Colour (APHA)	10 APHA	< 10
Chloride (Cl-)	1 ppb	< 1
Phosphate (PO ₄ ³⁻)	1 ppb	< 1
Sulphate (SO ₄ ²⁻)	1 ppb	< 1

Dr. B. McKelvey
 QA/QC Manager

Release Date: September 15, 2009
Expiry Date: September 15, 2012

Product Integrity:

Based on extensive testing results, SEASTAR CHEMICALS INC have found our products, unopened and sealed, maintain the certified integrity, or product quality, for a minimum of three years under the following conditions:

- Stored at room temperature, maximum range 15°C (59°F) to 25°C (77°F).
- Minimum exposure to light.
- For limited time, storage/transport temperature range 5°C (41°F) to 35°C (95°F)

Upon opening the product, the product's integrity will depend on proper handling and exposure to contaminants. The product has been bottled under CLASS 100 clean room conditions, to maintain the certified quality it should be used under these conditions. Furthermore to reduce trace metal contamination, the inner pack of plastic bags and bottle should be opened under CLASS 100 particle conditions to maintain the integrity of the product. The use of plastic gloves, hair net and a clean room suit is also advised.

Safety:

PRIOR to opening or storing this product be sure to consult the Material Safety Data Sheet (MSDS) Section 7 Handling and Storage to ensure safe storage and handling with regards to this hazardous material. This information must be understood prior to its use or storage.

SAFETY HANDLING NOTES: Consult your MSDS, PRIOR to handling these materials. Use proper safety apparel according to the recommendations of the MSDS. Exposure controls and personal protection should include: a properly functioning fume hood, protection for eyes (safety glasses), hands (chemically compatible gloves), feet (chemically compatible boots) and exposed skin (splash protection and a chemically compatible apron). All of these items must conform to local/regional/national regulatory requirements.

Dr. B. McKelvey
QA/QC Manager

10005 McDonald Park Road, Sidney, BC, Canada V8L 3S8
phone: (250) 655-5880 fax: (250) 655-5888
toll free: 1 (800) 663-2330 (within Canada & U.S. only)
Email: seastar.technicalsupport@seastarchemicals.com
Web: www.seastarchemicals.com