## CERTIFICATE OF ANALYSIS

# **BASELINE®** Water

1A		PRODUCT NUMBER: S020901								LOT NUMBER: 9216010							
3 Li < 0.01 11 Na < 5	Nitric Acid / 2% Hydrogen Peroxide. Operations are conducted under Class 100 or better clean-room conditions.  For volatile elements (indicated by *), the samples are diluted then directly injected into the ICP-MS. Values below  13 AI								7A								
		3B	4B	5B	6B	7B		8		1B	2B		20 0	00 4-*	04 0-1		
19 K	<b>20</b> Ca < 10	<b>21 Sc</b> < 0.01	<b>22</b> Ti < 0.02	23 V < 0.01	<b>24</b> Cr < 0.01	<b>25</b> Mn < 0.01	<b>26</b> Fe < 0.5	27 Co < 0.02	28 Ni < 0.2	< 0.05	30 Zn < 0.5	<b>31 Ga</b> < 0.01	<b>32 Ge</b> < 0.1	33 As* < 0.1	<b>34 Se*</b> < 0.1		
<b>37 Rb</b> < 0.01	38 Sr < 0.02	<b>39 Y</b> < 0.01	<b>40</b> Zr < 0.01	<b>41 Nb</b> < 0.01	<b>42</b> Mo < 0.1		<b>44</b> Ru < 0.01	45 Rh < 0.01	<b>46</b> Pd < 0.01	<b>47 Ag</b> < 0.01	48 Cd < 1	<b>49</b> In < 0.01	<b>50</b> Sn < 0.05		<b>52</b> Te < 0.02		
<b>55 Cs</b> < 0.01	<b>56 Ba</b> < 0.01	<b>57</b> La < 0.01	<b>72</b> Hf < 0.01	<b>73</b> Ta < 0.01	<b>74 W</b> < 0.5	<b>75</b> Re < 0.01			<b>78 Pt</b> < 0.01	<b>79 Au</b> < 0.1	<b>80 Hg*</b> < 10	<b>81 TI</b> < 0.01	<b>82 Pb</b> < 0.01	<b>83</b> Bi < 0.01			
				14						7							

#### ALL VALUES ARE REPORTED IN PARTS PER TRILLION (PPT)

KEY (1) (2) (2) (3) (4)

(1) Atomic Number

(2) Elemental Symbol (3) Concentration (mean in ppt)

(4) 1 Standard Deviation (N=3)

58 Ce	59 Pr	60 Nd	62 Sm	63 Eu	64 Gd	65 Tb	66 Dy	67 Ho	68 Er	69 Tm	70 Yb	71 Lu
< 0.01	< 0.01	< 0.01	< 0.01	< 0.01	< 0.01	< 0.01	< 0.01	< 0.01	< 0.01	< 0.01	< 0.01	< 0.01
90 Th	, ,	92 U										
< 0.01		< 0.01										
	L											



Additional	Maximum	Actual
Tests	Specification	Value
Colour (APHA)	10 APHA	< 10
Chloride (CI-)	1 ppb	< 1
Phosphate (PO <sub>4</sub> <sup>3</sup> -)	1 ppb	< 1
Sulphate (SO <sub>4</sub> <sup>2-</sup> )	1 ppb	< 1

Greg Henson QA & RA Manager Release Date: January 29, 2016 Expiry Date: January 29, 2019



## **Product Integrity:**

Based on extensive testing results, SEASTAR CHEMICALS INC have found our products, unopened and sealed, maintain the certified integrity, or product quality, for a minimum of three years under the following conditions:

- Stored at room temperature, maximum range 15°C (59°F) to 25°C (77°F).
- Minimum exposure to light.
- For limited time, storage/transport temperature range 5°C (41°F) to 35°C (95°F)

Upon opening the product, the product's integrity will depend on proper handling and exposure to contaminants. The product has been bottled under CLASS 100 clean room conditions, to maintain the certified quality it should be used under these conditions. Furthermore to reduce trace metal contamination, the inner pack of plastic bags and bottle should be opened under CLASS 100 particle conditions to maintain the integrity of the product. The use of plastic gloves, hair net and a clean room suit is also advised.

### Safety:

PRIOR to opening or storing this product be sure to consult the Material Safety Data Sheet (MSDS) Section 7 Handling and Storage to ensure safe storage and handling with regards to this hazardous material. This information must be understood prior to its use or storage.

SAFETY HANDLING NOTES: Consult your MSDS, PRIOR to handling these materials. Use proper safety apparel according to the recommendations of the MSDS. Exposure controls and personal protection should include: a properly functioning fume hood, protection for eyes (safety glasses), hands (chemically compatible gloves), feet (chemically compatible boots) and exposed skin (splash protection and a chemically compatible apron). All of these items must conform to local/regional/national regulatory requirements.

Greg Henson

Brey denon-

QA & RA Manager

10005 McDonald Park Road, Sidney, BC, Canada V8L 5Y2

phone: (250) 655-5880 fax: (250) 655-5888

toll free: 1 (800) 663-2330 (within Canada & U.S. only) Email: <a href="mailto:seastar.technicalsupport@seastarchemicals.com">seastar.technicalsupport@seastarchemicals.com</a>

Web: www.seastarchemicals.com