

# C E R T I F I C A T E   O F   A N A L Y S I S

## BASELINE<sup>®</sup> Bromine

**PRODUCT NUMBER: S021201**

**LOT NUMBER: 12209050**

1A		PRODUCT NUMBER: S021201										LOT NUMBER: 12209050														
2A		Elements are determined by magnetic sector ICP-MS using sample pre-concentration. The results are an average of three aliquots subsampled from three samples representative of the lot. The samples are slowly evaporated to dryness, the resulting residue is reconstituted in a small volume of 2% SEASTAR <sup>™</sup> BASELINE <sup>®</sup> Nitric Acid. Operations are conducted under Class 100 particle or better clean-room conditions. Values below 3 times the standard deviation of the blank are shown with "<", no blank value is subtracted.										3A		4A		5A		6A		7A						
3 Li < 10	4 Be < 10											5 B < 100														
11 Na < 20	12 Mg < 20											13 Al < 50														
		3B	4B	5B	6B	7B	8		1B	2B																
19 K < 20	20 Ca < 50	21 Sc < 10	22 Ti < 10	23 V < 10	24 Cr < 10	25 Mn < 10	26 Fe < 50	27 Co < 10	28 Ni < 10	29 Cu < 10	30 Zn < 20	31 Ga < 10														
37 Rb < 10	38 Sr < 10	39 Y < 10	40 Zr < 10	41 Nb < 100	42 Mo < 100		44 Ru < 10	45 Rh < 100	46 Pd < 100		48 Cd < 10	49 In < 10	50 Sn < 200	51 Sb < 50	52 Te < 10											
55 Cs < 10	56 Ba < 10	57 La < 10	72 Hf < 5		74 W < 10	75 Re < 10			78 Pt < 100			81 Tl < 10	82 Pb < 10	83 Bi < 10												

**ALL VALUES ARE REPORTED IN PARTS PER TRILLION (PPT)**

KEY	(1)	Atomic Number	58 Ce < 10	59 Pr < 10	60 Nd < 10		62 Sm < 10	63 Eu < 10	64 Gd < 10	65 Tb < 10	66 Dy < 10	67 Ho < 10	68 Er < 10	69 Tm < 10	70 Yb < 10	71 Lu < 10
	(1) (2)	Elemental Symbol														
	(3)	Concentration (mean in ppt)														
	(4)	1 Standard Deviation (N=3)	90 Th < 10		92 U < 10											

**Release Date: December 22, 2009**

**Expiry Date: December 22, 2012**



  
 Dr. B. McKelvey  
 QA/QC Manager

## Product Integrity:

Based on extensive testing results, SEASTAR CHEMICALS INC have found our products, unopened and sealed, maintain the certified integrity, or product quality, for a minimum of three years under the following conditions:

- Stored at room temperature, maximum range 15°C (59°F) to 25°C (77°F).
- Minimum exposure to light.
- For limited time, storage/transport temperature range 5°C (41°F) to 35°C (95°F)

Upon opening the product, the product's integrity will depend on proper handling and exposure to contaminants. The product has been bottled under CLASS 100 clean room conditions, to maintain the certified quality it should be used under these conditions. Furthermore to reduce trace metal contamination, the inner pack of plastic bags and bottle should be opened under CLASS 100 particle conditions to maintain the integrity of the product. The use of plastic gloves, hair net and a clean room suit is also advised.

## Safety:

PRIOR to opening or storing this product be sure to consult the Material Safety Data Sheet (MSDS) Section 7 Handling and Storage to ensure safe storage and handling with regards to this hazardous material. This information must be understood prior to its use or storage.

SAFETY HANDLING NOTES: Consult your MSDS, PRIOR to handling these materials. Use proper safety apparel according to the recommendations of the MSDS. Exposure controls and personal protection should include: a properly functioning fume hood, protection for eyes (safety glasses), hands (chemically compatible gloves), feet (chemically compatible boots) and exposed skin (splash protection and a chemically compatible apron). All of these items must conform to local/regional/national regulatory requirements.



Dr. B. McKelvey  
QA/QC Manager

10005 McDonald Park Road, Sidney, BC, Canada V8L 5Y2  
phone: (250) 655-5880 fax: (250) 655-5888  
toll free: 1 (800) 663-2330 (within Canada & U.S. only)  
Email: [seastar.technicalsupport@seastarchemicals.com](mailto:seastar.technicalsupport@seastarchemicals.com)