



## Certificate of Analysis

<b>Product Number:</b>	<b>S011001</b>	<b>CAS Number:</b>	7722-84-1
<b>Product Description:</b>	Hydrogen Peroxide Solution	<b>Molecular Weight:</b>	34.01 g/mol
<b>Product Grade:</b>	Instrument Quality	<b>Molecular Formula:</b>	H <sub>2</sub> O <sub>2</sub>
<b>Lot Number:</b>	17122124	<b>Density:</b>	1.11 g/mL
<b>Release Date:</b>	01/21/2026 (mm/dd/yyyy)	<b>Molarity:</b>	10 moles/litre
<b>Expiration Date:</b>	01/21/2028 (mm/dd/yyyy)	<b>Normality:</b>	10 moles/litre

### Analytical Data

Analyte	Specification	Actual Value	Analyte	Specification	Actual Value
Assay (H <sub>2</sub> O <sub>2</sub> )	30 - 32% w/w	31% w/w	Mercury (Hg)	0.1 ppb	< 0.02 ppb
Colour	10 APHA	< 10 APHA	Molybdenum (Mo)	0.1 ppb	< 0.1 ppb
Chloride (Cl <sup>-</sup> )	300 ppb	< 300 ppb	Neodymium (Nd)	0.1 ppb	< 0.1 ppb
Aluminum (Al)	1 ppb	< 0.5 ppb	Nickel (Ni)	0.5 ppb	< 0.2 ppb
Antimony (Sb)	0.5 ppb	< 0.1 ppb	Niobium (Nb)	0.1 ppb	< 0.1 ppb
Arsenic (As)	0.5 ppb	< 0.1 ppb	Palladium (Pd)	0.5 ppb	< 0.1 ppb
Barium (Ba)	0.1 ppb	< 0.1 ppb	Platinum (Pt)	0.5 ppb	< 0.1 ppb
Beryllium (Be)	0.1 ppb	< 0.1 ppb	Potassium (K)	1 ppb	< 0.1 ppb
Bismuth (Bi)	0.1 ppb	< 0.1 ppb	Praseodymium (Pr)	0.1 ppb	< 0.1 ppb
Boron (B)	1 ppb	< 0.5 ppb	Rhenium (Re)	0.1 ppb	< 0.1 ppb
Cadmium (Cd)	0.5 ppb	< 0.1 ppb	Rhodium (Rh)	0.5 ppb	< 0.1 ppb
Calcium (Ca)	1 ppb	< 0.5 ppb	Rubidium (Rb)	0.1 ppb	< 0.1 ppb
Cerium (Ce)	0.1 ppb	< 0.1 ppb	Ruthenium (Ru)	0.5 ppb	< 0.1 ppb
Cesium (Cs)	0.1 ppb	< 0.1 ppb	Samarium (Sm)	0.1 ppb	< 0.1 ppb
Chromium (Cr)	1 ppb	< 0.5 ppb	Scandium (Sc)	0.1 ppb	< 0.1 ppb
Cobalt (Co)	0.5 ppb	< 0.1 ppb	Selenium (Se)	1 ppb	< 0.2 ppb
Copper (Cu)	0.5 ppb	< 0.1 ppb	Silver (Ag)	0.1 ppb	< 0.1 ppb
Dysprosium (Dy)	0.1 ppb	< 0.1 ppb	Sodium (Na)	1 ppb	< 0.5 ppb
Erbium (Er)	0.1 ppb	< 0.1 ppb	Strontium (Sr)	0.1 ppb	< 0.1 ppb
Europium (Eu)	0.1 ppb	< 0.1 ppb	Tantalum (Ta)	0.1 ppb	< 0.1 ppb
Gadolinium (Gd)	0.1 ppb	< 0.1 ppb	Tellurium (Te)	0.1 ppb	< 0.1 ppb
Gallium (Ga)	0.1 ppb	< 0.1 ppb	Terbium (Tb)	0.1 ppb	< 0.1 ppb
Germanium (Ge)	0.1 ppb	< 0.1 ppb	Thallium (Tl)	0.1 ppb	< 0.1 ppb
Gold (Au)	0.1 ppb	< 0.1 ppb	Thorium (Th)	0.1 ppb	< 0.1 ppb
Hafnium (Hf)	0.1 ppb	< 0.1 ppb	Thulium (Tm)	0.1 ppb	< 0.1 ppb
Holmium (Ho)	0.1 ppb	< 0.1 ppb	Tin (Sn)	0.5 ppb	< 0.1 ppb
Indium (In)	0.1 ppb	< 0.1 ppb	Titanium (Ti)	0.5 ppb	< 0.1 ppb
Iron (Fe)	1 ppb	< 0.5 ppb	Tungsten (W)	0.1 ppb	< 0.1 ppb
Lanthanum (La)	0.1 ppb	< 0.1 ppb	Uranium (U)	0.1 ppb	< 0.1 ppb
Lead (Pb)	0.1 ppb	< 0.1 ppb	Vanadium (V)	0.5 ppb	< 0.1 ppb
Lithium (Li)	0.1 ppb	< 0.1 ppb	Ytterbium (Yb)	0.1 ppb	< 0.1 ppb
Lutetium (Lu)	0.1 ppb	< 0.1 ppb	Yttrium (Y)	0.1 ppb	< 0.1 ppb
Magnesium (Mg)	1 ppb	< 0.2 ppb	Zinc (Zn)	0.5 ppb	< 0.5 ppb
Manganese (Mn)	0.1 ppb	< 0.1 ppb	Zirconium (Zr)	0.1 ppb	< 0.1 ppb

The hydrogen peroxide solution used to make this product had a chloride content of < 300 ppb.

Greg Henson  
QA & RA Manager

For terms and conditions of use, please see page 2.

## Terms and Conditions of Use

### Safety Guidelines:

PRIOR to opening or storing this product be sure to consult the Safety Data Sheet (SDS) to ensure safe storage and handling with regards to this hazardous material. This information must be read and understood prior to use or storage.

SAFETY HANDLING NOTES: Consult the SDS PRIOR to handling this product. Use proper safety apparel according to the recommendations of the SDS. Exposure controls and personal protection should include: a properly functioning fume hood, protection for eyes (safety glasses), hands (chemically compatible gloves), feet (chemically compatible boots), and exposed skin (splash protection and a chemically compatible apron). All of these items must conform to local/regional/national regulatory requirements.

### SEASTAR™'s Product Integrity Guidelines:

We have found our products, unopened and sealed, maintain the certified integrity, or product quality, for their stated certification period under the following conditions:

- Store at room temperature, maximum range 15°C (59°F) to 25°C (77°F).
- Avoid exposure to sunlight or ultraviolet light sources.
- Open in a 'particle free' environment. SEASTAR recommends a HEPA or ULPA particle filtered trace metal clean room. Open product should be handled under Class 100 or ISO 5 clean room or better conditions.

Once opened, product integrity will depend on proper handling and exposure to contaminants. To reduce trace metal contamination, the inner pack of plastic bags and bottle should be opened under Class 100 or ISO 5 clean room or better conditions to maintain the integrity of the product. The use of plastic gloves, hair net and a clean room suit is also advised.

**For SEASTAR™'s Product Expiration Policy and Product Permeation FAQ, please see our website.**

### Notes:

Reported density, molarity and normality values reflect published literature and are characteristic of the product's assay range. If you require an accurate density, molarity, or normality for the product that you have purchased, you will have to perform the measurement. Bottles within a given lot have small assay variations.

### Definitions:

- **Actual value:** the measured value in a particular lot analysis.
- **Analyte:** the substance being measured.
- **Specification:** the maximum certified value of an analyte, unless otherwise specified.
- **Unit(s):** **ppm** – part per million or µg (microgram) of analyte per gram of solution.  
**ppb** – part per billion or ng (nanogram) of analyte per gram of solution.  
**ppt** – part per trillion or pg (picogram) of analyte per gram of solution.



Greg Henson  
QA & RA Manager